



# Certificate of Compliance

**Certificate:** 70061842

**Master Contract:** 180737

**Project:** 70120319

**Date Issued:** 2017-02-14

**Issued to:** K-Patents OY  
PL 77  
Vantaa, 01511  
FINLAND

**Attention:** Veikko Kahkonen

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



**Issued by:** Sam Zaffino  
Sam Zaffino

## **PRODUCTS**

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT-Certified to US Standards

Model: MI Multichannel User Interface; rated 100-240Vac, 50/60Hz, 60VA (AC version) or 24Vdc, 60W (DC version); two sets of relay contacts rated 3.0A max at 150Vac, all other inputs and outputs SELV, Ambient 0 to 50°C, Pollution degree 4, Installation Category II, Enclosure Type 4X.

## **APPLICABLE REQUIREMENTS**

- |   |  |
|---|--|
| CAN/CSA-C22.2 No. 61010-1-12                | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements  |
| ANSI/UL Std. No. 61010-1 (3 <sup>rd</sup> ) | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |



## *Supplement to Certificate of Compliance*

**Certificate:** 70061842

**Master Contract:** 180737

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
70120319	2017-02-14	Update to 70061842 to cover alt construction
70061842	2016-08-09	Acceptance of CB Report for all electrical testing for North America on Multichannel User Interface includes Type 4, 4X and “door closed” Type 6, 6P, 12 and 13.