



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

VAISALA INC
10 Gill Street, Suite D
Woburn, MA 01801
Dan Soave Phone: 781 537 1082

CALIBRATION

Valid To: December 31, 2019

Certificate Number: 2083.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Pressure	(60 to 1100) hPa	0.066 hPa	Fluke DHI PPC4

II. Thermodynamics

Parameter/Equipment	Range	CMC ² (±)	Comments
Relative Humidity	(10 to 20) % RH (> 20 to 40) % RH (> 40 to 95) % RH	0.42 % RH 0.60 % RH 0.60 % RH	Thunder Scientific 2500
	0 % RH	0.10 % RH	Vaisala DMP348 dewpoint transmitter

Parameter/Equipment	Range	CMC ² (±)	Comments
Temperature	(-80 to 200) °C	0.054 °C	Fluke Hart Scientific 2560 w/5614 or 5615 probe
	(-110 to 110) °C	0.020 °C	Fluke Hart Scientific 2560 w/5626 PRT
	(0 to 15) °C	0.34 °C	Thunder Scientific 2500
	(> 15 to 25) °C	0.13 °C	
(> 25 to 40) °C	0.19 °C		
Dewpoint	(> -60 to -10) °C	0.73 °C	Thunder Scientific 3900
	(> -10 to 10) °C	0.80 °C	MBW 373LX
	(-80 to -60) °C	0.90 °C	

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.





Accredited Laboratory

A2LA has accredited

VAISALA INC.

Woburn, MA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).



Presented this 8th day of February 2018.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 2083.01
Valid to December 31, 2019

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.