



## DL1016 Temperature Logger

Please choose one attribute from each numbered category and add the corresponding mark to the grid on top of the table.

As a result configuration code for the Vaisala device is formed. Example of order code with typical settings:

DL1016	V	A	1	N	N	N	N	N
--------	---	---	---	---	---	---	---	---

**Selections in bold are included in the prices of the basic versions.**

*Selections in italic are available at an extra price.*

Vaisala Data Logger Configuration			DL1016	1	2	3	4	5	6	7	8
<b>1. Validation Type</b>	<i>Validation</i>		V								
	<b>Non Validation</b>		S								
<b>2. Channel Function</b>	<i>Temperature Stainless Steel</i>	<i>3 m (10 ft.) Probe</i>		A							
<b>Length Ext 1</b>	<i>Temperature Stainless Steel</i>	<i>7.6 m (25 ft.) Probe</i>		B							
	<i>Temperature Water Proof</i>	<i>3 m (10 ft.) Probe</i>		C							
	<i>Temperature Water Proof</i>	<i>7.6 m (25 ft.) Probe</i>		D							
	<i>Temperature Stainless Steel</i>	<i>60 Feet</i>		E							
<b>3. Channel Function</b>	<b>None</b>							N			
<b>Length Ext 2</b>	<i>Temperature Stainless Steel</i>	<i>3 m (10 ft.) Probe</i>		1							
	<i>Temperature Stainless Steel</i>	<i>7.6 m (25 ft.) Probe</i>		2							
	<i>Temperature Water Proof</i>	<i>3 m (10 ft.) Probe</i>		3							
	<i>Temperature Water Proof</i>	<i>7.6 m (25 ft.) Probe</i>		4							
	<b>Boolean</b>	<i>7.6 m (25 ft.) Probe</i>		5							
	<b>Boolean</b>	<b>No Cables</b>		6							
	<i>Temperature Stainless Steel</i>	<i>60 Feet</i>		7							
<b>4. Additional Temperature Calibration Points*</b>	<b>None</b>							N			
	<i>1 Additional Point**</i>	<i>Specify Additional Point here</i>		A							
	<i>2 Additional Points**</i>	<i>Specify Additional Points here</i>		B							
<b>5. Preset Calibration Notification Interval</b>	<b>Recommended</b>	<b>12 Months</b>						N			
	<b>Customer Specified</b>	<b>3 - 120 Months</b>						1			
									N		
										N	
											N

\*Additional 1 week delivery

**Additional Calibration Points	Temperature Range
Stainless Steel and Water Proof Probes	-90 to +50 Deg C