



HMP115T Temperature Probe

Please choose one attribute from each numbered category and add the corresponding mark to the grid on top of the table.
 As a result configuration code for the Vaisala device is formed. Example of order code with typical settings:

HMP115T | **C** | **1** | **B** | **1** | **A**

Selections in bold are included in the prices of the basic versions.
Selections in italic are available at an extra price.

				1	2	3	4	5
Vaisala Temperature Probe Configuration		HMP115T		C		B	1	A
1. Probe Type and Standard Calibration	3-point T	0°C, 20°C, 40°C (32°F, 68°F, 104°F)		C				
2. Sensor Protection	Plastic Grid Protection	For quick response time in fast changing conditions		1				
	<i>Membrane Filter</i>	<i>For quick response time in dusty environments</i>		<i>2</i>				
	<i>Sintered PTFE filter</i>	<i>For quick response time in the harshest conditions with e.g. water droplets</i>		<i>3</i>				
						B		
							1	
								A

Tick the box to order Accredited Calibration (HELACCRCAL)

Tick the box to order Custom Calibration* (HELSPECIALCAL)
 Specify Custom Calibration Point(s) here

* Range for Custom Calibration	
Temperature	-40 to +60 °C (-40 to +140 °F)

